



UNITED STATES DEPARTMENT OF COMMERCE
Patent and Trademark Office

Address: COMMISSIONER OF PATENTS AND TRADEMARKS
Washington, D.C. 20231

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.
-----------------	-------------	----------------------	---------------------

09/666,000 09/20/00 ROFFERS

R 15-XT-4847

EXAMINER

MMC2/0913

JOHN F NETHERY
MCANDREWS HELD & MALLOY LTD
34TH FLOOR
500 W MADISON STREET
CHICAGO IL 60661

CHURCH, C

ART UNIT

PAPER NUMBER

2882

DATE MAILED:

09/13/01

Please find below and/or attached an Office communication concerning this application or proceeding.

Commissioner of Patents and Trademarks

Office Action Summary

Application No.

Applicant(s)

Examiner

Group Art Unit

—The MAILING DATE of this communication appears on the cover sheet beneath the correspondence address—

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If the period for reply specified above is less than thirty (30) days, a reply within the statutory minimum of thirty (30) days will be considered timely.
- If NO period for reply is specified above, such period shall, by default, expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133).

Status

- ☐ Responsive to communication(s) filed on _____.
- ☐ This action is **FINAL**.
- ☐ Since this application is in condition for allowance except for formal matters, **prosecution as to the merits is closed** in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11; 453 O.G. 213.

Disposition of Claims

- ☒ Claim(s) 1-20 is/are pending in the application.
- Of the above claim(s) _____ is/are withdrawn from consideration.
- ☐ Claim(s) _____ is/are allowed.
- ☒ Claim(s) 1-20 is/are rejected.
- ☐ Claim(s) _____ is/are objected to.
- ☐ Claim(s) _____ are subject to restriction or election requirement.

Application Papers

- ☐ See the attached Notice of Draftsperson's Patent Drawing Review, PTO-948.
- ☐ The proposed drawing correction, filed on _____ is ☐ approved ☐ disapproved.
- ☐ The drawing(s) filed on _____ is/are objected to by the Examiner.
- ☐ The specification is objected to by the Examiner.
- ☐ The oath or declaration is objected to by the Examiner.

Priority under 35 U.S.C. § 119 (a)-(d)

- ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d).
 - ☐ All ☐ Some* ☐ None of the CERTIFIED copies of the priority documents have been:
 - ☐ received.
 - ☐ received in Application No. (Series Code/Serial Number) _____.
 - ☐ received in this national stage application from the International Bureau (PCT Rule 1.7.2(a)).

*Certified copies not received: _____

Attachment(s)

- ☐ Information Disclosure Statement(s), PTO-1449, Paper No(s). _____
- ☒ Notice of Reference(s) Cited, PTO-892
- ☐ Notice of Draftsperson's Patent Drawing Review, PTO-948
- ☐ Interview Summary, PTO-413
- ☐ Notice of Informal Patent Application, PTO-152
- ☐ Other _____

Office Action Summary

Serial No. 666,000
Art Unit 2882

-2-

The drawings are objected to under 37 C.F.R. § 1.83(a). The drawings must show every feature of the invention specified in the claims. Therefore, the means for supplying the ion collection voltage must be shown or the feature canceled from the claims. No new matter should be entered.

The following is a quotation of the first paragraph of 35 U.S.C. § 112:

The specification shall contain a written description of the invention, and of the manner and process of making and using it, in such full, clear, concise, and exact terms as to enable any person skilled in the art to which it pertains, or with which it is most nearly connected, to make and use the same and shall set forth the best mode contemplated by the inventor of carrying out his invention.

The specification is objected to under 35 U.S.C. § 112, first paragraph, as failing to provide an adequate written description and failing to provide an enabling disclosure. There is no teaching of where or how the ion collection voltage is applied in addition to the "focus voltage" from supply 124 as there is only one grid.

Claims 1-20 are rejected under 35 U.S.C. § 112, first paragraph, for the reasons set forth in the objection to the specification.

Claims 1-20 rejected under 35 U.S.C. § 112, second paragraph, as being indefinite for failing to particularly point out and distinctly claim the subject matter which applicant regards as the invention. The claims are incomplete for failing to recite

Serial No. 666,000
Art Unit 2882

-3-

essential elements of the invention such as a cathode, anode and grid. No electron focus means are defined in addition to the ion collection grid. The meanings of "calibrating the x-ray tube" in claim 11 and "an x-ray detector positioned to receive the electron beam" in claim 14 are unclear.

The following is a quotation of 35 U.S.C. § 103 which forms the basis for all obviousness rejections set forth in this Office action:

A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

Subject matter developed by another person, which qualifies as prior art only under subsection (f) or (g) of section 102 of this title, shall not preclude patentability under this section where the subject matter and the claimed invention were, at the time the invention was made, owned by the same person or subject to an obligation of assignment to the same person.

Claims 1-20 are rejected under 35 U.S.C. § 103 as being unpatentable over Baptist (6259765 PCT published 12/98}. Baptist teaches an x-ray system comprising an x-ray tube having a vacuum housing 2, window 14, cathode 6, target anode 10, ion suppression grid 17, power supplies 28 and 29 and detector 32. See lines 35-40 of column 4 and 59-63 of column 6. The claims are rejected to the extent they are understood.

Any inquiry concerning this communication should be directed

Serial No. 666,000
Art Unit 2882

-4-

to Examiner Church at telephone number (703) 308-4861.

Craig E Church

CRAIG E. CHURCH
Senior Examiner
ART UNIT 2882